

## CERTIFICATE OF ANALYSIS

Client: VanaH CO., Ltd.  
1648 Araya, Fujiyoshida-shi, Yamanashi 403-0006, Japan

Sample name: VanaH (product) 2L Lot 2012. 2. 14 L11A/F

Received date: February 14, 2012

This is to certify that the following result(s) have been obtained from our analysis on the above-mentioned sample(s) submitted by the client.

### Test Result(s)

Test Item	Result	QL	N	M
Radioactive nuclide	.....	.....	1	
Cesium-137	Not detected	20 Bq/kg	2	
Cesium-134	Not detected	20 Bq/kg	2	
Iodine-131	Not detected	20 Bq/kg	2	

QL: Quantitation limit N: Notes M: Method


<Radioactive nuclide> Test facility: Tama Laboratory, Japan Food Research Laboratories (6-11-10 Nagayama, Tama-shi, Tokyo, Japan)

### Notes

1:  $\gamma$ -ray spectrometer (germanium semiconductor detector).

2: Measuring period, 1000 seconds.



  
Noriko Imaizumi  
Principal Investigator

Feb. 20, 2012  
Date

## CERTIFICATE OF ANALYSIS

Client: VanaH CO., Ltd.  
1648 Araya, Fujiyoshida-shi, Yamanashi 403-0006, Japan

Sample name: VanaH (raw water) Lot 2012.2.14

Received date: February 14, 2012

This is to certify that the following result(s) have been obtained from our analysis on the above-mentioned sample(s) submitted by the client.

### Test Result(s)

Test Item	Result	QL	N	M
Radioactive nuclide	.....	.....	1	
Cesium-137	Not detected	20 Bq/kg	2	
Cesium-134	Not detected	20 Bq/kg	2	
Iodine-131	Not detected	20 Bq/kg	2	

QL: Quantitation limit N: Notes M: Method

<Radioactive nuclide> Test facility: Tama Laboratory, Japan Food Research Laboratories (6-11-10 Nagayama, Tama-shi, Tokyo, Japan)

### Notes

1:  $\gamma$ -ray spectrometer (germanium semiconductor detector).

2: Measuring period, 1000 seconds.



Noriko Imaizumi  
Principal Investigator



Date